



## Certificate of Analysis

Product Description	Brain Heart Infusion Broth		Cat. No.	EX-CM917
Specification	250g/bottle		Batch No.	240818
Basic Requirements	Appearance	Light yellow powder	Solution color	Light yellow
Basic Requirements	Odor	Non-pungent odor	Gel Strength	-
	pH (25°C)	7.3	Loss On Drying	<5%
	Microbial Contamination	-	Clarity	Good
Microbiological Test	Test Standard	LBT/ZY-10-040-2024		
	Control Microorganism	Inoculum CFU	Evaluation Standard	Test Result
	<i>Escherichia coli</i> ATCC 25922	10-100	Well grown (Turbidity 2)	Well grown (Turbidity 2)
	<i>Pseudomonas aeruginosa</i> ATCC 25619	10-100	Well grown (Turbidity 2)	Well grown (Turbidity 2)
	<i>Enterococcus faecalis</i> ATCC 29212	10-100	Well grown (Turbidity 2)	Well grown (Turbidity 2)
	<i>Beta hemolytic streptococcus</i> CMCC (B) 32210	10-100	Well grown (Turbidity 2)	Well grown (Turbidity 2)
Remark	The culture conditions is 36°C/24h.			
Test Result	Qualified			
Operator & Date	Jingjing Jing 21/08/2024	Reviewer & Date	Zhiran Yu 21/08/2024	Auditor & Date Pei Liu 21/08/2024

