



### Certificate of Analysis

<b>Product Description</b>	Sampling Bag	<b>Cat. No.</b>	EX-GT2212	
<b>Specification</b>	100 packs	<b>Batch No.</b>	231204	
<b>Exp Date</b>	3 Years	<b>Test Standard</b>	LBT/ZY-QC-05-006-2019	
<b>Testing Item</b>		<b>Evaluation Standard</b>		<b>Test Result</b>
<b>Appearance</b>		The seal contains a metal wire		Conform
<b>Color</b>		White		Conform
<b>Character</b>		Evenness, Transparency		Conform
<b>Odor</b>		Non-pungent odor		Conform
<b>Impurity</b>		No visible impurities		Conform
<b>Length</b>		30±0.3 cm		30.1cm
<b>Width</b>		17.5±0.3 cm		17.6cm
<b>Sealing</b>		No leakage		Conform
<b>Microbial Contamination</b>		Aseptic growth		Conform
<b>Test Result</b>	<b>Qualified</b>			
<b>Operator &amp; Date:</b>	Jin Zhang 08/12/2023	<b>Verifier &amp; Date:</b>	Bing Li 08/12/2023	<b>Auditor &amp; Date:</b> Liu 08/12/2023

