



### Certificate of Analysis

<b>Product Description</b>	1%TTC Solution		<b>Cat. No.</b>	EX-P-79B-10mL
<b>Specification</b>	10mL /bottle		<b>Batch No.</b>	240228
<b>Basic Requirements</b>	<b>Appearance</b>	Liquid	<b>Color</b>	Colourless
	<b>Odor</b>	Non-pungent odor	<b>Gel Strength</b>	-
	<b>pH (25°C)</b>	-	<b>Loss On Drying</b>	-
	<b>Microbial Contamination</b>	No bacterial growth	<b>Clarity</b>	Good
<b>Microbiological Test</b>	<b>Control Microorganism</b>	<b>Evaluation Standard</b>	<b>Test Result</b>	<b>Characteristic Reaction</b>
	<i>Enterococcus faecalis</i> ATCC 29212	Good	Good	Red colonies
	<i>Escherichia coli</i> ATCC 25922	Inhibited	Inhibited	-
	<i>Enterobacter aerogenes</i> ATCC 13048	Inhibited	Inhibited	-
<b>Remarks</b>	<p>1.Add 1mL EX-P-79B-10mL 1%TTC solution(1 vial) to 100mL EX-CM165A basis.</p> <p>2.The inoculation amount:ATCC29212:10~100CFU/mL;Others:1,000~10,000CFU/mL.</p> <p>3.The incubation time:36°C/48h (The time that result is obtained)</p>			
<b>Test Result</b>	<b>Qualified</b>			
<b>Operator &amp; Date</b>	Bing Li 04/03/2024	<b>Reviewer &amp; Date</b>	Zhiran Yu 04/03/2024	<b>Auditor &amp; Date</b> 04/03/2024

